




3M™ Petrifilm™ Aqua- Yeast and Mold Count Plates

Product Information

MEDIUM:	Petrifilm™ Aqua Yeast and Mold Count Plates (6408/6413)													
ISO/BSI CERTIFICATE OF REGISTRATION NUMBER:	3M Food Safety is certified to ISO-9001:2008, FM 14552 Manufactured at Brookings, South Dakota, U.S.A.													
DATE OF EXPIRATION/ LOT NUMBER	Expiry and lot number indicated on each package. Lot number indicated on each plate.													
FORMULATION:	SABHI nutrients, water soluble gel, bromo-chloro-indolyphosphate substrate													
METHOD OF PREPARATION:	Nutrients and gels coated onto film. For use with membrane filtration, place filtered sample onto plate and hydrate with one mL diluent. See product instructions for detailed information.													
CONTAMINATION CHECK:	Minimum 80 plates per batch tested Incubated at 25°C for 5 days Columbia sequential sampling plan													
EFFICACY CHECK:	Organisms tested include: <table><thead><tr><th>Organism</th><th>Acceptable Batch</th></tr></thead><tbody><tr><td><i>Enterococcus faecalis</i></td><td>No Growth</td></tr><tr><td><i>Acinetobacter</i></td><td>No Growth</td></tr><tr><td><i>Aspergillus niger</i></td><td rowspan="4">Counts not lower than 3 standard deviations below the count on Rose Bengal agar with chlorotetracycline for all molds and yeasts.</td></tr><tr><td><i>Cladosporium cladosporioides</i></td></tr><tr><td><i>Hansenula anomala</i></td></tr><tr><td><i>Saccharomyces cerevisiae</i></td></tr><tr><td><i>Candida albicans</i></td><td></td></tr></tbody></table>	Organism	Acceptable Batch	<i>Enterococcus faecalis</i>	No Growth	<i>Acinetobacter</i>	No Growth	<i>Aspergillus niger</i>	Counts not lower than 3 standard deviations below the count on Rose Bengal agar with chlorotetracycline for all molds and yeasts.	<i>Cladosporium cladosporioides</i>	<i>Hansenula anomala</i>	<i>Saccharomyces cerevisiae</i>	<i>Candida albicans</i>	
Organism	Acceptable Batch													
<i>Enterococcus faecalis</i>	No Growth													
<i>Acinetobacter</i>	No Growth													
<i>Aspergillus niger</i>	Counts not lower than 3 standard deviations below the count on Rose Bengal agar with chlorotetracycline for all molds and yeasts.													
<i>Cladosporium cladosporioides</i>														
<i>Hansenula anomala</i>														
<i>Saccharomyces cerevisiae</i>														
<i>Candida albicans</i>														
PACKAGING:	Pack size: 50 plates per foil pouch Film grade: Plastic foil laminate Seal integrity check: Delamination seal integrity test performed													
MEASUREMENT SYSTEMS CALIBRATION AND TRACEABILITY:	Incubator temperature. 3M internal calibration. Minimum calibration once per year for all equipment.													
MEDIA QUALITY STATEMENT:	Quality assurance certificate included in package													
SHELF LIFE:	18 months from date of manufacture													
STORAGE CONDITIONS:	Store at temperatures less than or equal to 8°C													
SIGNED:	 M.M. Henrickson Quality Assurance													